Certificate
PCR Performance Tested quality



For the use of SARSTEDT consumables in nucleic acid analytics

PCR Performance Tested items fulfil the following purity criteria certified by an independent laboratory:

- ✓ DNA-free
- ✓ DNase/RNase-free
- ✓ PCR inhibitor-free

We guarantee that the following limit values are met:

Human DNA <0.5 pg/µl
Bacterial DNA <0.02 pg/µl
DNase $<1x10^{-5}$ U/µl
RNase $<1x10^{-9}$ Kunitz units/µl
Real Time PCR <0.5 cycles Ct value shifting compared to control

SARSTEDT AG & Co. KG

Central Quality Assurance

Technical modifications reserved

GDO_20_005_0100_NA_NF