



Product Specification
Sample tube, Citrate 3.2%, 2 ml, cap violet, (LxØ):
66 x 11.5 mm, with print

Page 1



Product description

Order number	47.410
Product description	Sample tube, nominal volume: 2 ml, cap violet, preparation: Trisodium citrate solution, screw cap, (LxØ): 66 x 11.5 mm, skirted conical base, with print, 100 piece(s)/bag

Product characteristics

Type of collection	venous
Type of preparation	Trisodium citrate solution
Colour code	EU
Label/ Print	with print
Colour of print/label	violet
Graduation	yes
Closure type	screw cap
Suitable for	Sediplus®
Base shape	skirted conical base

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2022-07-06

This document was prepared by EDP support and is valid without signature.



SARSTEDT



Product Specification
Sample tube, Citrate 3.2%, 2 ml, cap violet, (LxØ):
66 x 11.5 mm, with print

Page 2

Size

Sample volume	2 ml
Diameter	11.5 mm
Length of product	66 mm
Length including cap	67 mm
Length excluding cap	66 mm

Material & colours

Colour of product	transparent
Closure material	High Density Polyethylene (HD-PE)
Colour of cap	violet
Tube material	Polypropylene (PP)

Purity & certification

Satisfies the requirement	IATA, ADR
Product category	In-vitro diagnostic, CE
CE certified	CE
Batched	yes

Packaging

Minimum order qty.	500
Type of inner packaging	bag
Piece(s) / inner box	100
Piece(s) / case	500
Piece(s) / pallet	48000
Depth of case	348 mm
Width of case	278 mm
Height of case	140 mm
Case volume	0.0135 cbm
Weight of product	0.0027 kg
Weight of case	1.64 kg
EAN of inner box	4038917081443
EAN case	4038917000536

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2022-07-06

This document was prepared by EDP support and is valid without signature.



SARSTEDT