



Product description

Order number	20.1308
Product description	Microvette® 100 Serum, capillary blood collection, preparation: Clotting activator, nominal volume: 300 µl, (LxØ) with cap: 47.6 x 10.8 mm, cap white, flat base, inner receptacle: conical

Product characteristics

Type of collection	capillary
Type of preparation	Clotting activator
Colour code	EU
Label/ Print	with print
Colour of print/label	white
Cap	quadruple-threaded cap
Application	capillary blood collection
Base shape	flat base
Recommended centrifugation	10000 x g
Centrifugation time	5 Min
Shape of inner tube	conical

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.





Product Specification
Microvette® 100 Serum, 300 µl, cap white, flat base

Page 2

Size

Sample volume	300 µl
Diameter	10,8 mm
Length including cap	47,6 mm
Length excluding cap	46,6 mm

Material & colours

Product material	Polypropylene (PP)
Colour of product	transparent
Cap material	High Density Polyethylene (HD-PE)
Colour of cap	white

Purity & certification

Product category	IVD product
CE certified	CE
Batched	yes

Packaging

Minimum order qty.	1000
Type of inner packaging	inner box
Piece(s) / inner box	100
Piece(s) / case	1000
Piece(s) / pallet	121000
Depth of box	139 mm
Width of box	108 mm
Height of box	66 mm
Depth of case	338 mm
Width of case	228 mm
Height of case	160 mm
Case volume	0,0123 cbm
Weight of product	0,0019 kg
Weight of case	2,42 kg
EAN of inner box	4038917086134
EAN case	4038917004718

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2021-12-03

This document was prepared by EDP support and is valid without signature.



SARSTEDT