



Product Specification

**Screw cap tube, 3.5 ml, (LxØ): 92 x 13 mm,
conical false bottom, rounded tube bottom, PP,
cap assembled, 100 piece(s)/bag**

Page 1



Product description

Order number	62.617
Product description	Screw cap tube, working volume: 3.5 ml, (LxØ): 92 x 13 mm, conical false bottom, rounded tube bottom, transparent, material: PP, with graduations, cap assembled, 100 piece(s)/bag, 1,000 piece(s)/carton

Product characteristics

Graduation	yes
Base shape	conical false bottom, rounded tube bottom
Volume of work	3,5 ml

Size

Diameter	13 mm
Length of product	92 mm

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2024-03-29

This document was prepared by EDP
support and is valid without signature.



SARSTEDT



Product Specification

**Screw cap tube, 3.5 ml, (LxØ): 92 x 13 mm,
conical false bottom, rounded tube bottom, PP,
cap assembled, 100 piece(s)/bag**

Page 2

Material & colours

Product material	Polypropylene (PP)
Colour of product	transparent

Purity & certification

Satisfies the requirement	IATA, ADR
Product category	In vitro diagnostic device, CE
CE certified	CE
Batched	yes

Packaging

Minimum order qty.	1000
Type of smallest subpackaging	bag
Piece(s) / inner box	100
Piece (s) / outer case	1000
Piece(s) / pallet	36000
Depth of case	458 mm
Width of case	408 mm
Height of case	205 mm
Case volume	0.0383 cbm
Weight of product	0.0036 kg
Weight of case	4.34 kg
EAN of inner box	4038917082921
EAN case	4038917079150

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2024-03-29

This document was prepared by EDP support and is valid without signature.



SARSTEDT