

Product Specification Micro sample tube Lithium heparin LH, 1.3 ml, push cap, ISO



Product description

Order number Product description

41.1503.105

Micro sample tube Lithium heparin LH, venous blood collection, preparation: Lithium heparin, nominal volume: 1.3 ml, $(Lx\emptyset)$ with cap: 44 x 10.8 mm, push cap, cap: green, colour code ISO, with plastic label, label/print: green, skirted conical base, 100 piece(s)/bag

Product characteristics

Type of collection Type of preparation Preparation concentration Colour code Label/ Print Colour of print/label Graduation Cap Closure type Application Base shape venous Lithium heparin 25 IU/ml blood ISO with plastic label green yes flat cap push cap venous blood collection skirted conical base

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This document was prepared by EDP support and is valid without signature.





Volume of work

1,3 ml

| Recommended centrifugation Centrifugation time | 2000 x g 5 Min |
|---|--|
| Size | |
| Sample volume Diameter Length of product Length including cap Length excluding cap | 1.3 ml 10.8 mm 44 mm 44 mm 40 mm |
| Material & colours | |
| Colour of product Closure material Colour of cap Tube material | transparent Low Density Polyethylene (LD-PE) green Polypropylene (PP) |
| Purity & certification | |
| Product category CE certified Batched | In vitrodiagnostic device, CE CE yes |
| Packaging | |
| Minimum order qty. Type of smallest subpackaging | 1000 bag |
| Piece(s) / inner box Piece (s) / outer case Piece(s) / pallet Depth of case Width of case Height of case Case volume Weight of product Weight of case EAN of inner box EAN case | 100 1000 96000 348 mm 278 mm 140 mm 0.0135 cbm 0.0016 kg 1.88 kg 4038917129541 4038917022286 |
| L/114 0000 | |

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This document was prepared by EDP support and is valid without signature.

