

## Product Specification S-Monovette<sup>®</sup> Serum Gel CAT, 4.7 ml, cap brown, (LxØ): 75 x 15 mm, with plastic label



#### **Product description**

| Order number        |
|---------------------|
| Product description |

#### 03.1524

S-Monovette<sup>®</sup> Serum Gel CAT, preparation: Clotting activator/gel, 4.7 ml, membrane screw cap, cap brown, colour code EU/ISO, (LxØ) without cap: 75 x 15 mm, with plastic label, label/print: transparent/brown, 50 piece(s)/case, sterile

#### **Product characteristics**

| Type of collection         | venous                 |
|----------------------------|------------------------|
| Type of preparation        | Clotting activator/gel |
| Colour code                | EU/ISO                 |
| Label/ Print               | with plastic label     |
| Colour of print/label      | transparent/brown      |
| Graduation                 | fill mark              |
| Сар                        | membrane screw cap     |
| Closure type               | screw cap              |
| Recommended centrifugation | 2500 x g               |
| Centrifugation time        | 10 Min                 |
| Liquid-tight               | yes                    |

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This document was prepared by EDP support and is valid without signature.





Page 2

#### Connection

membrane screw cap

### Size

| Sample volume            | 4.7 ml |
|--------------------------|--------|
| Diameter                 | 15 mm  |
| Length including cap     | 91 mm  |
| Length excluding cap     | 75 mm  |
| Length including cap and | 112 mm |
| plunger                  |        |

## Material & colours

| transparent                       |
|-----------------------------------|
| High Density Polyethylene (HD-PE) |
| brown                             |
| rubber                            |
| Polypropylene (PP)                |
| High Density Polyethylene (HD-PE) |
| Polystyrene (PS)                  |
|                                   |

## **Purity & certification**

| Satisfies the requirement | IATA, ADR                     |
|---------------------------|-------------------------------|
| Product category          | In vitrodiagnostic device, CE |
| CE certified              | CE                            |
| Purity standard           | sterile                       |
| Sterilisation             | Electron irradiation          |
| Batched                   | yes                           |
| Batched                   | yes                           |

# Packaging

| Minimum order qty.     | 500    |
|------------------------|--------|
| Type of smallest       | case   |
| subpackaging           |        |
| Piece(s) / inner box   | 50     |
| Piece (s) / outer case | 500    |
| Piece(s) / pallet      | 26000  |
| Depth of box           | 189 mm |
| Width of box           | 117 mm |
| Height of box          | 82 mm  |

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.





### Product Specification S-Monovette<sup>®</sup> Serum Gel CAT, 4.7 ml, cap brown, (LxØ): 75 x 15 mm, with plastic label

Page 3

| Depth of case     |
|-------------------|
| Width of case     |
| Height of case    |
| Case volume       |
| Weight of product |
| Weight of case    |
| EAN of inner box  |
| EAN case          |

440 mm 400 mm 129 mm 0.0227 cbm 0.0074 kg 4.52 kg 4038917086868 4038917005395

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This document was prepared by EDP support and is valid without signature.

