

Product Specification S-Monovette[®] Lithium heparin gel LH, 2.6 ml, cap orange, (LxØ): 65 x 13 mm, with plastic label

Page 1



Product description

| Order number |
|---------------------|
| Product description |

04.1907

S-Monovette[®] Lithium heparin gel LH, preparation: Lithium heparin gel, 2.6 ml, membrane screw cap, cap orange, colour code EU, (LxØ) without cap: 65 x 13 mm, with plastic label, label/print: transparent/brown, 50 piece(s)/case, sterile

Product characteristics

| Type of collection | venous |
|----------------------------|---------------------|
| Type of preparation | Lithium heparin gel |
| Preparation concentration | 25 IU/ml blood |
| Colour code | EU |
| Label/ Print | with plastic label |
| Colour of print/label | transparent/brown |
| Graduation | fill mark |
| Сар | membrane screw cap |
| Closure type | screw cap |
| Recommended centrifugation | 2500 x g |
| Centrifugation temperature | 20 |

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This document was prepared by EDP support and is valid without signature.





Product Specification S-Monovette[®] Lithium heparin gel LH, 2.6 ml, cap orange, (LxØ): 65 x 13 mm, with plastic label

Page 2

| Centrifugation time | |
|---------------------|--|
| Liquid-tight | |
| Connection | |

15 Min yes membrane screw cap

Size

| Sample volume | 2.6 ml |
|--------------------------|--------|
| Diameter | 13 mm |
| Length including cap | 81 mm |
| Length excluding cap | 65 mm |
| Length including cap and | 102 mm |
| plunger | |

Material & colours

| Colour of product | transparent |
|-------------------------|-----------------------------------|
| Closure material | High Density Polyethylene (HD-PE) |
| Colour of cap | orange |
| Membrane material | rubber |
| Tube material | Polypropylene (PP) |
| Piston material | High Density Polyethylene (HD-PE) |
| Piston plunger material | Polystyrene (PS) |
| | |

IATA, ADR

Electron irradiation

CE

yes

sterile

In vitrodiagnostic device, CE

Purity & certification

Satisfies the requirement Product category CE certified Purity standard Sterilisation Batched

Packaging

| 500 |
|--------|
| case |
| |
| 50 |
| 500 |
| 42000 |
| 171 mm |
| |

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This document was prepared by EDP support and is valid without signature.





Width of box Height of box Depth of case Width of case Height of case Case volume Weight of product Weight of case EAN of inner box EAN case 105 mm 73 mm 388 mm 358 mm 125 mm 0.0174 cbm 0.0048 kg 3.12 kg 4038917088015 4038917006170

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

Date of issue: 2024-05-19

This document was prepared by EDP support and is valid without signature.

