




Product Specification
**Serological pipette, plugged, 2 ml, sterile,
non-pyrogenic/endotoxin-free, non-cytotoxic, 1
piece(s)/blister**

Page 1



Product description

Order number	86.1252.001
Product description	Serological pipette, plugged, max. working volume: 2 ml, coding colour: green, material: PS, sterile, non-pyrogenic/endotoxin-free, non-cytotoxic, 1 piece(s)/blister

Product characteristics

Colour of print/label	green
Graduation	yes
Design	plugged
Volume of work	2 ml
Coding colour	green

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2021-10-17

This document was prepared by EDP
support and is valid without signature.



SARSTEDT



Product Specification
**Serological pipette, plugged, 2 ml, sterile,
non-pyrogenic/endotoxin-free, non-cytotoxic, 1
piece(s)/blister**

Page 2

Size

Diameter 6 mm
Length of product 270 mm

Material & colours

Product material Polystyrene (PS)
Colour of product transparent

Purity & certification

Product category no medical device | IVD
Purity standard sterile, non-pyrogenic/endotoxin-free, non-cytotoxic
Sterilisation Gamma irradiation
Pyrogen-free yes
Batched yes

Packaging

Minimum order qty. 1000
Type of inner packaging blister
Piece(s) / inner box 100
Piece(s) / case 1000
Piece(s) / pallet 54000
Case volume 0,0323 cbm
Weight of product 0,0044 kg
Weight of case 5,03 kg
EAN of inner box 4038917127196
EAN case 4038917021081

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2021-10-17

This document was prepared by EDP support and is valid without signature.



SARSTEDT