



Product description

Order number	72.1981.010
Product description	PCR plate half skirt, 96 well, white, Low Profile, 100 µl, PCR Performance Tested, material: PP, optimised for qPCR, 10 piece(s)/minigrip bag

Product characteristics

Suitable for qPCR	yes
Base shape	conical
Construction of the plate edges	half skirt
High or low profile	Low Profile
Number of wells	96
Volume of work	5 µl - 100 µl
Centrifugation max (RCF)	3000 x g
Centrifugation temperature	22
Centrifugation time	5 Min
Autoclavable (max. °C)	121 °C

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.





Product Specification
PCR plate half skirt, 96 well, white, Low Profile,
100 µl, PCR Performance Tested, PP

Page 2

Size

Sample volume	0.1 ml
Max. volume	100 µl
Diameter	5.48 mm
Width of product	85.4 mm
Height of product	17.37 mm
Length of product	127.8 mm

Material & colours

Product material	Polypropylene (PP)
Colour of product	white

Purity & certification

Product category	no medical device IVD
Purity standard	PCR Performance Tested
Batched	yes

Packaging

Minimum order qty.	100
Type of smallest subpackaging	minigrip bag
Piece(s) / inner box	50
Piece (s) / outer case	100
Piece(s) / pallet	9800
Depth of box	270 mm
Width of box	230 mm
Height of box	96 mm
Depth of case	282 mm
Width of case	210 mm
Height of case	247 mm
Case volume	0.0146 cbm
Weight of product	0.0154 kg
Weight of case	1.89 kg
EAN of inner box	4038917573672
EAN case	4038917573665

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2024-03-28

This document was prepared by EDP support and is valid without signature.



SARSTEDT