

Product Specification PCR strip of 4, 200 μl, Biosphere[®] plus, transparent, PP, flat cap





Product description

•	
Order number	72.990
Product description	PCR strip of 4, up to 200 μ l, High Profile, Biosphere [®] plus,
	transparent, material: PP, with anti-contamination shield, flat

cap, cap attached, suitable for qPCR, 50 piece(s)/bag

Product characteristics

Graduation	no
Сар	flat cap
Closure type	snap-on lid
Suitable for qPCR	yes
Base shape	conical
High or low profile	High Profile
Number of wells	4
Volume of work	10 µl - 0,2 ml
Centrifugation max (RCF)	8000 x g
Centrifugation temperature	22
Centrifugation time	5 Min
Autoclavable (max. °C)	121 °C

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This document was prepared by EDP support and is valid without signature.





Size

Sample volume Max. volume Diameter Width of product Height of product Length of product	0.2 ml 200 µl 5.35 mm 10.6 mm 20.65 mm 37.2 mm
Material & colours	
Product material Colour of product	Polypropylene (PP) transparent
Purity & certification	
Product category CE certified Purity standard Sterilisation Pyrogen-free Batched	In vitrodiagnostic device, CE CE Biosphere [®] plus Gas sterilisation yes yes
Packaging	
Packaging Minimum order qty. Type of smallest subpackaging	400 bag
Minimum order qty. Type of smallest	
Minimum order qty. Type of smallest subpackaging	bag
Minimum order qty. Type of smallest subpackaging Piece(s) / inner box Piece (s) / outer case	bag 200 400

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.





Product Specification PCR strip of 4, 200 μl, Biosphere[®] plus, transparent, PP, flat cap

Page 3

EAN case

4038917030045

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

Date of issue: 2024-05-05

This document was prepared by EDP support and is valid without signature.

