



### **Product description**

Order number	55.476.061
Product description	Tube, working volume: 5 ml, (LxØ): 75 x 12 mm, material: PS, round base, transparent, push cap, suitable for Vitek and Vitek 2 from bioMérieux, 500 piece(s)/StackPack

### **Product characteristics**

Graduation	no
Closure type	push cap
Suitable for	Vitek and Vitek 2 from bioMérieux
Base shape	round base
Volume of work	5 ml
Centrifugation max (RCF)	4000 x g

### **Size**

Diameter	12 mm
Length of product	75 mm

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.





**Product Specification**  
**Tube, 5 ml, (LxØ): 75 x 12 mm, PS**

**Page 2**

**Material & colours**

Product material	Polystyrene (PS)
Colour of product	transparent

**Purity & certification**

Satisfies the requirement	sealed at 0.6 bar/10 min
Product category	In vitro diagnostic device, CE
CE certified	CE
Batched	yes

**Packaging**

Minimum order qty.	2000
Type of smallest subpackaging	StackPack
Piece(s) / inner box	500
Piece (s) / outer case	2000
Piece(s) / pallet	120000
Depth of box	387 mm
Width of box	77 mm
Height of box	176 mm
Depth of case	398 mm
Width of case	358 mm
Height of case	175 mm
Case volume	0.0249 cbm
Weight of product	0.0028 kg
Weight of case	6.2 kg
EAN of inner box	4038917107631
EAN case	4038917011839

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

Date of issue:  
2024-04-23

This document was prepared by EDP support and is valid without signature.



**SARSTEDT**