



Product description

Order number	72.730.406
Product description	Screw cap micro tube, working volume: 0.5 ml, skirted conical base, with knurling, transparent, cap: natural, cap assembled, with printed writing space, PCR Performance Tested, 100 piece(s)/bag

Product characteristics

Label/ Print	with print
Colour of print/label	white
Graduation	no
Cap	standard screw cap
Closure type	screw cap
Knurls	yes
Base shape	skirted conical base
Volume of work	0,5 ml
Centrifugation max (RCF)	20000 x g

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.





Product Specification
Screw cap micro tube, 0.5 ml, PCR Performance
Tested

Page 2

Size

Diameter	10.8 mm
Length of product	44 mm

Material & colours

Product material	Polypropylene (PP)
Colour of product	transparent
Closure material	Polypropylene (PP)
Colour of cap	natural
O-ring material	Ethylene propylene diene rubber (EPDM)

Purity & certification

Satisfies the requirement	IATA
Product category	In vitrodiagnostic device, CE
CE certified	CE
Purity standard	PCR Performance Tested
Batched	yes

Packaging

Minimum order qty.	1000
Type of smallest subpackaging	bag
Piece(s) / inner box	500
Piece (s) / outer case	1000
Piece(s) / pallet	110000
Depth of box	211 mm
Width of box	191 mm
Height of box	141 mm
Depth of case	398 mm
Width of case	223 mm
Height of case	160 mm
Case volume	0.0142 cbm
Weight of product	0.0015 kg
Weight of case	2.04 kg
EAN of inner box	4038917223980
EAN case	4038917223973

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2024-03-29

This document was prepared by EDP support and is valid without signature.



SARSTEDT