



Product description

Order number	41.1378.005
Product description	Micro sample tube Serum Gel CAT, venous blood collection, preparation: Clotting activator/gel, nominal volume: 1.1 ml, (LxØ) with cap: 47 x 10.8 mm, screw cap, cap: brown, colour code EU/ISO, with paper label, label/print: brown, skirted conical base, 100 piece(s)/bag

Product characteristics

Type of collection	venous
Type of preparation	Clotting activator/gel
Colour code	EU/ISO
Label/ Print	with paper label
Colour of print/label	brown
Graduation	yes
Closure type	screw cap
Application	venous blood collection
Base shape	skirted conical base
Volume of work	1,1 ml
Recommended centrifugation	10000 x g

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.





Product Specification
Micro sample tube Serum Gel CAT, 1.1 ml, screw cap, EU/ISO

Page 2

Centrifugation time 5 Min

Size

Sample volume 1.1 ml
Diameter 10.8 mm
Length of product 47 mm
Length including cap 47 mm
Length excluding cap 45 mm

Material & colours

Colour of product transparent
Closure material Polypropylene (PP)
Colour of cap brown
Tube material Polypropylene (PP)

Purity & certification

Satisfies the requirement IATA, ADR
Product category In vitrodiagnostic device, CE
CE certified CE
Batched yes

Packaging

Minimum order qty. 1000
Type of smallest subpackaging bag
Piece(s) / inner box 100
Piece (s) / outer case 1000
Piece(s) / pallet 96000
Depth of case 348 mm
Width of case 278 mm
Height of case 140 mm
Case volume 0.0135 cbm
Weight of product 0.0019 kg
Weight of case 2.17 kg
EAN of inner box 4038917127967
EAN case 4038917021500

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

Date of issue:
2024-05-08

This document was prepared by EDP support and is valid without signature.



SARSTEDT