



Product Specification
**Urine drainage system, 2 l, with anti-reflux valve,
with T drainage valve, sterile**

Page 1



Product description

| | |
|---------------------|---|
| Order number | 74.5485.005 |
| Product description | Urine drainage system, 2 l, with anti-reflux valve, with T drainage valve, sterile, tube length: 100 cm, 1 piece(s)/blister |
| PZN number | 16356199 |

Product characteristics

| | |
|-----------------------|-----------------------|
| Colour of print/label | blue |
| Design | with T drainage valve |
| Application | Urine drainage |
| Volume of work | 2 l |
| Anti-reflux valve | yes |

Size

| | |
|------------------|---------|
| Sample volume | 2 l |
| Max. volume | 2 l |
| Diameter of tube | 10.1 mm |
| Length of tube | 100 cm |

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.





Product Specification
**Urine drainage system, 2 l, with anti-reflux valve,
with T drainage valve, sterile**

Page 2

Material & colours

| | |
|-------------------|-------------------------|
| Product material | Polyvinylchloride (PVC) |
| Colour of product | transparent |

Purity & certification

| | |
|------------------|-------------------|
| Product category | Medical device |
| CE certified | CE supplier |
| Purity standard | sterile |
| Sterilisation | Gas sterilisation |
| Batched | yes |

Packaging

| | |
|-------------------------------|---------------|
| Minimum order qty. | 50 |
| Type of smallest subpackaging | blister |
| Piece(s) / inner box | 1 |
| Piece (s) / outer case | 50 |
| Piece(s) / pallet | 2400 |
| Depth of case | 395 mm |
| Width of case | 295 mm |
| Height of case | 290 mm |
| Case volume | 0.0338 cbm |
| Weight of product | 0.0844 kg |
| Weight of case | 4.62 kg |
| EAN of inner box | 6955824024418 |
| EAN case | 6955824024036 |

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

Date of issue:
2024-05-02

This document was prepared by EDP support and is valid without signature.



SARSTEDT